



# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R85-1/2008-NL1-14.01  
Project number 14200253  
Page 1 of 2

Issuing authority NMI Certin B.V.  
Person responsible: C. Oosterman

Applicant and Manufacturer Enraf B.V.  
Delftechpark 39  
2628 XJ Delft  
The Netherlands

Identification of the certified type Automatic level gauge for measuring the level of liquid in storage tanks, models Smartradar Flexline XP and Smartradar Flexline HP, with indicating device, with the antennas F06, F08, W06, H04, S06, S08, S10 and S12.

Types : Level gauge SmartRadar Flexline XP  
Level gauge SmartRadar Flexline HP  
Indicating device SmartView

Characteristics See page 2

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R85-1/2008 (E):** Metrological and technical requirements

**R85-2/2008 (E):** Test methods

**R85-3/2008 (E):** Test Report format

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMI Certin B.V., OIML Issuing Authority NL1**  
4 December 2014

  
C. Oosterman  
Head Certification Board

NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMI Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see [www.nmi.nl](http://www.nmi.nl)).





# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R85-1/2008-NL1-14.01  
Project number 14200253  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- Number R85/1998-NL1-07.02 that includes 100 pages.
- Number CPC/9200376 that includes 20 pages.
- Number NMI-10200994 that includes 15 pages.
- Number NMI-12200691 that includes 13 pages.
- Number NMI-13200623 that includes 14 pages.
- Number NMI-14200253-1 that includes 21 pages.

## Identification of the certified pattern

Automatic level gauge for measuring the level of liquid in storage tanks, models SmartRadar Flexline XP and SmartRadar Flexline HP, with the antennas F06, F08, W06, H04, S06, S08, S10 and S12.

Antenna type	Minimum product level	Maximum product level	Minimum and maximum values for liquid pressure, for liquid temperature and for liquid properties.	Minimum and maximum values for vapour pressure, for vapour temperature and for vapour properties.
F06, F08, W06	20 m below the antenna	1 m below the antenna	The manufacturer shall specify these values for each application.	The manufacturer shall specify these values for each application.
H04, S06, S08, S10, S12	21 m below the antenna	1 m below the antenna	The manufacturer shall specify these values for each application.	The manufacturer shall specify these values for each application.